



tel. +420 545 555 111, fax +420 545 222 728, www.cmi.cz

EC-TYPE EXAMINATION CERTIFICATE

Number: TCM 143/14 - 5230

Page 1 from 11 pages

In accordance:

with Directive 2004/22/EC of the European Parliament and of the Council as amended

implemented in Czech Republic by Government Order No. 464/2005 Coll. as amended that

lays down technical requirements on measuring instruments.

Manufacturer:

Metreg Technologies GmbH

Tränkeweg 9

D-15517 Fürstenwalde

Germany

For:

volume conversion device

type: MECflexM

MPE 0.5 %

mechanical environment class: M2

electromagnetic environment class: E2

climatic environments limits: - 25°C...+ 70 °C

Valid until:

28 October 2024

Document No:

0511-CS-A069-14

Description:

Essential characteristics, approved conditions and special conditions, if any, are described in

this certificate.

Date of issue:

29 October 2014

Certificate approved by:



RNDr. Pavel Klenovský